

Penlon MRI Product Range

7. Diamond MRI Laryngoscope Range

NOTE

The Penlon Diamond MRI Laryngoscope Range is certified for use in magnetic field strengths up to 3.0 tesla, in a spatial gradient field of up to 720 gauss/cm.

Description

Handle: Standard size

Handle: Short

Lithium Battery, 2-pack

Note:

Lithium batteries are classed as Dangerous Goods Class 9 (Ref. UN 3090 legislation).

Batteries for export markets are subject to air freight surcharges to facilitate compliance with UN 3090.



BLade: American Macintosh 0

BLade: American Macintosh 1

BLade: American Macintosh 2

BLade: American Macintosh 3

BLade: American Macintosh 3.5

Blade: American Macintosh 4



English Macintosh 1

Blade: English Macintosh 2

Blade: English Macintosh 3

Blade: English Macintosh 4



Blade: Miller 0

Blade: Miller 1

BLade: Miller 2

BLade: Miller 3

Blade: Miller 4



Blade: Wisconsin 2

Blade: Wisconsin 3

Blade: Wisconsin 4



Xenon Light Bulb, 5-pack

8. MRI Facility Equipment

MRI Flowmeter, Oxygen, single tube, direct probe, nipple outlet



Bubble Humidifier - autoclavable

